

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0137_25648118/1**

Product	Ethyl acetate, GC-MS	Batch	25648118
AC0137		Quality release date	14/03/2025
		Expiry date	03/2028

Analysis	Batch value	Specifications
Assay (G.C.)	99,98 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Colour (Hazen)	3	≤ 10
Residue on evaporation	1,1 ppm	≤ 3
Water (K.F.)	0,014 %	≤ 0,05
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0

Suitable for residue analysis.

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Avilés Palou
Laboratory Technician