

**Scharlab S.L.**

Crtra. Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0148_24928005/1**

Product	Ethyl acetate, for GC residue analysis	Batch	24928005
AC0148		Quality release date	21/06/2024
		Expiry date	06/2027
Analysis	Batch value	Specifications	
Assay (G.C.)	99,97 %	≥ 99,8	
Identity (IR-Spectrum)	passes test	passes test	
Density (20°/4°)	0,9007 g/mL	0,900 - 0,902	
Residue on evaporation	0,0001 %	≤ 0,0001	
Water (K.F.)	0,016 %	≤ 0,02	

Suitable for organohalogenated pesticide and dioxins, furans and PCBs residue analysis.
ECD, from 1,2,4-trichlorobenzene to decachlorobiphenyl, no peaks are obtained greater than 3 pg/ml as lindane.
No peaks are obtained in vicinity of 2,4,5-trichlorobiphenyl.

For laboratory use only