

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AC0293\_23921422/1**

Product	Acetone, GC-MS	Batch	23921422
<b>AC0293</b>		Quality release date	31/08/2023
		Expiry date	08/2026
<b>Analysis</b>	<b>Batch value</b>	<b>Specifications</b>	
Assay (G.C.)	99,98 %	≥ 99,8	
Colour (Hazen)	1	≤ 10	
Identity (IR-Spectrum)	passes test	passes test	
Residue on evaporation	1,2 ppm	≤ 3	
Water (K.F.)	0,164 %	≤ 0,2	
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0	

Suitable for residue analysis.

For laboratory use only.