

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0293_24146722/1**

Product	Acetone, GC-MS	Batch	24146722
AC0293		Quality release date	31/01/2024
		Expiry date	01/2027
Analysis	Batch value	Specifications	
Assay (G.C.)	99,99 %	≥ 99,8	
Colour (Hazen)	4	≤ 10	
Identity (IR-Spectrum)	passes test	passes test	
Residue on evaporation	1,0 ppm	≤ 3	
Water (K.F.)	0,1956 %	≤ 0,2	
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0	

Suitable for residue analysis.

For laboratory use only.