

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0293_25380822/1**

Product	Acetone, GC-MS	Batch	25380822
AC0293		Quality release date	09/12/2024
		Expiry date	12/2027
Analysis	Batch value	Specifications	
Assay (G.C.)	99,99 %	≥ 99,8	
Colour (Hazen)	3	≤ 10	
Identity (IR-Spectrum)	passes test	passes test	
Residue on evaporation	1,0 ppm	≤ 3	
Water (K.F.)	0,16 %	≤ 0,2	
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0	

Suitable for residue analysis.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Avilés Palou
Laboratory Technician