

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS**

Product	Acetone, for GC residue analysis	Batch	18785805
<b>AC0308</b>		Quality release date	09/04/2018
		Expiry date	03/2021
<b>Analysis</b>	<b>Batch value</b>	<b>Specifications</b>	
Assay (G.C.)	99,98 %	≥ 99,8	
Identity (IR-Spectrum)	passes test	passes test	
Density (20°/4°)	0,7908 g/mL	0,787 - 0,791	
Residue on evaporation	0,0001 %	≤ 0,0001	
Water (K.F.)	0,19 %	≤ 0,2	

Suitable for organohalogenated pesticide and dioxins, furans and PCBs residue analysis.

ECD, from 1,2,4-trichlorobenzene to decachlorobiphenyl, no peaks are obtained greater than 3 pg/ml as lindane.

No peaks are obtained in vicinity of 2,4,5-trichlorobiphenyl.