



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

AC0310_22992009/1

Product Acetone, Multisolvant® HPLC grade ACS ISO UV-VIS

Batch 22992009

AC0310

Quality release date 18/07/2022

Expiry date 07/2027

Analysis	Batch value	Specifications
Assay (G.C.)	99,97 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7909 g/mL	0,787 - 0,791
Appearance of solution	passes test	passes test
Colour (Hazen)	1	≤ 10
Solubility in water	passes test	passes test
Insoluble in water	passes test	passes test
Acidity	0,00009 meq/g	≤ 0,0002
Alkalinity	0,00002 meq/g	≤ 0,0002
Chlorides (Cl)	< 0,00001 %	≤ 0,00001
Nitrates (NO3)	< 0,00001 %	≤ 0,00001
Phosphates (as PO4)	< 0,00001 %	≤ 0,00001
Sulfates (SO4)	< 0,00001 %	≤ 0,00001
Aluminium (Al)	< 0,1 ppm	≤ 0,1
Antimony (Sb)	< 0,02 ppm	≤ 0,02
Arsenic (As)	< 0,02 ppm	≤ 0,02
Barium (Ba)	< 0,01 ppm	≤ 0,01
Beryllium (Be)	< 0,02 ppm	≤ 0,02
Bismuth (Bi)	< 0,02 ppm	≤ 0,02
Boron (B)	< 0,02 ppm	≤ 0,02
Cadmium (Cd)	< 0,01 ppm	≤ 0,01
Calcium (Ca)	< 0,3 ppm	≤ 0,3
Chromium (Cr)	< 0,02 ppm	≤ 0,02
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,01 ppm	≤ 0,01
Gallium (Ga)	< 0,02 ppm	≤ 0,02
Germanium (Ge)	< 0,02 ppm	≤ 0,02
Gold (Au)	< 0,02 ppm	≤ 0,02
Indium (In)	< 0,02 ppm	≤ 0,02
Iron (Fe)	< 0,02 ppm	≤ 0,02
Lead (Pb)	< 0,01 ppm	≤ 0,01
Lithium (Li)	< 0,05 ppm	≤ 0,05
Magnesium (Mg)	< 0,1 ppm	≤ 0,1
Manganese (Mn)	< 0,01 ppm	≤ 0,01
Molybdenum (Mo)	< 0,02 ppm	≤ 0,02
Nickel (Ni)	< 0,01 ppm	≤ 0,01
Silver (Ag)	< 0,02 ppm	≤ 0,02
Thallium (Tl)	< 0,02 ppm	≤ 0,02
Tin (Sn)	< 0,1 ppm	≤ 0,1
Titanium (Ti)	< 0,02 ppm	≤ 0,02
Vanadium (V)	< 0,02 ppm	≤ 0,02

Results marked with # are out of specifications.

This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Ricard Balladares Abella
Laboratory Technician



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

AC0310_22992009/1

Product Acetone, Multisolvant® HPLC grade ACS ISO UV-VIS

Batch 22992009

AC0310

Quality release date 18/07/2022

Expiry date 07/2027

Analysis	Batch value	Specifications
Zinc (Zn)	< 0,01 ppm	≤ 0,01
Zirconium (Zr)	< 0,02 ppm	≤ 0,02
Aldehydes (as HCHO)	< 0,002 %	≤ 0,002
Cyclohexane (G.C.)	< 0,1 %	≤ 0,1
Alcohol diacetone (G.C.)	0,0002 %	≤ 0,02
Ethanol (G.C.)	< 0,01 %	≤ 0,01
Methanol (G.C.)	0,0178 %	≤ 0,05
2-Propanol (G.C.)	0,0049 %	≤ 0,05
Reducing substances	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,0002
Water (K.F.)	0,2 %	≤ 0,2
Liquid chromatography suitability absorbance	passes test	passes test
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 330 nm	17,97 %	≥ 10
min. transmission at 335 nm	60,00 %	≥ 50
min. transmission at 339 nm	83,37 %	≥ 80
min. transmission at 342 nm	92,74 %	≥ 90
min. transmission at 350 nm	99,71 %	≥ 98
max. absorbance at 330 nm	0,745 AU	≤ 1,000
max. absorbance at 335 nm	0,222 AU	≤ 0,301
max. absorbance at 339 nm	0,079 AU	≤ 0,097
max. absorbance at 342 nm	0,033 AU	≤ 0,046
max. absorbance at 350 nm	0,001 AU	≤ 0,009

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Ricard Balladares Abella
Laboratory Technician