

CERTIFICATE OF ANALYSIS

AC0310_26142509/1

Product Acetone, Multisolvent® HPLC grade ACS ISO UV-VIS

Batch 26142509

AC0310

Quality release date 30/09/2025

Expiry date 09/2030

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7906 g/mL	0,787 - 0,791
Appearance of solution	passes test	passes test
Colour (Hazen)	4	≤ 10
Solubility in water	passes test	passes test
Insoluble in water	passes test	passes test
Acidity	0,00013 meq/g	≤ 0,0002
Alkalinity	0,00001 meq/g	≤ 0,0002
Chlorides (Cl)	< 0,00001 %	≤ 0,00001
Nitrates (NO ₃)	< 0,00001 %	≤ 0,00001
Phosphates (as PO ₄)	< 0,00001 %	≤ 0,00001
Sulfates (SO ₄)	< 0,00001 %	≤ 0,00001
Aluminium (Al)	< 0,1 ppm	≤ 0,1
Antimony (Sb)	< 0,02 ppm	≤ 0,02
Arsenic (As)	< 0,02 ppm	≤ 0,02
Barium (Ba)	< 0,01 ppm	≤ 0,01
Beryllium (Be)	< 0,02 ppm	≤ 0,02
Bismuth (Bi)	< 0,02 ppm	≤ 0,02
Boron (B)	< 0,02 ppm	≤ 0,02
Cadmium (Cd)	< 0,01 ppm	≤ 0,01
Calcium (Ca)	< 0,3 ppm	≤ 0,3
Chromium (Cr)	< 0,02 ppm	≤ 0,02
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,01 ppm	≤ 0,01
Gallium (Ga)	< 0,02 ppm	≤ 0,02
Germanium (Ge)	< 0,02 ppm	≤ 0,02
Gold (Au)	< 0,02 ppm	≤ 0,02
Indium (In)	< 0,02 ppm	≤ 0,02
Iron (Fe)	< 0,02 ppm	≤ 0,02
Lead (Pb)	< 0,01 ppm	≤ 0,01
Lithium (Li)	< 0,05 ppm	≤ 0,05
Magnesium (Mg)	< 0,1 ppm	≤ 0,1
Manganese (Mn)	< 0,01 ppm	≤ 0,01
Molybdenum (Mo)	< 0,02 ppm	≤ 0,02
Nickel (Ni)	< 0,01 ppm	≤ 0,01
Silver (Ag)	< 0,02 ppm	≤ 0,02
Thallium (Tl)	< 0,02 ppm	≤ 0,02
Tin (Sn)	< 0,1 ppm	≤ 0,1
Titanium (Ti)	< 0,02 ppm	≤ 0,02

Results marked with # are out of specifications.
 This certificate does not exempt the user from checking the results upon receipt of the goods.
 Any copy of our CoA may be obtained from our website at www.scharlab.com

 Marta Martinez Casals
 Laboratory Technician



**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0310_26142509/1**

Product Acetone, Multisolvent® HPLC grade ACS ISO UV-VIS

Batch 26142509

AC0310

Quality release date 30/09/2025

Expiry date 09/2030

Analysis	Batch value	Specifications
Vanadium (V)	< 0,02 ppm	≤ 0,02
Zinc (Zn)	< 0,01 ppm	≤ 0,01
Zirconium (Zr)	< 0,02 ppm	≤ 0,02
Aldehydes (as HCHO)	< 0,002 %	≤ 0,002
Cyclohexane (G.C.)	< 0,005 %	≤ 0,1
Alcohol diacetone (G.C.)	< 0,005 %	≤ 0,02
Ethanol (G.C.)	< 0,005 %	≤ 0,01
Methanol (G.C.)	0,015 %	≤ 0,05
2-Propanol (G.C.)	< 0,005 %	≤ 0,05
Reducing substances	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,0002
Water (K.F.)	0,20 %	≤ 0,2
Liquid chromatography suitability absorbance	passes test	passes test
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 330 nm	22,59 %	≥ 10
min. transmission at 335 nm	65,89 %	≥ 50
min. transmission at 339 nm	87,26 %	≥ 80
min. transmission at 342 nm	94,59 %	≥ 90
min. transmission at 350 nm	99,76 %	≥ 98
max. absorbance at 330 nm	0,646 AU	≤ 1,000
max. absorbance at 335 nm	0,181 AU	≤ 0,301
max. absorbance at 339 nm	0,059 AU	≤ 0,097
max. absorbance at 342 nm	0,024 AU	≤ 0,046
max. absorbance at 350 nm	0,001 AU	≤ 0,009

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician