

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0346_23893408/1**

Product Acetic acid glacial, HPLC grade

Batch 23893408

AC0346

Quality release date 07/06/2023

Expiry date 06/2028

Analysis	Batch value	Specifications
Assay (Acidimetric)	99,96 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	1,0494 g/mL	1,048 - 1,050
Residue on evaporation	0,0002 %	≤ 0,0002
Water (K.F.)	0,0504 %	≤ 0,2
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 260 nm	81,93 %	≥ 80
min. transmission at 270 nm	96,36 %	≥ 95
min. transmission at 280 nm	98,00 %	≥ 98
max. absorbance at 260 nm	0,087 AU	≤ 0,097
max. absorbance at 270 nm	0,016 AU	≤ 0,022
max. absorbance at 280 nm	0,009 AU	≤ 0,009

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only