

**Scharlab S.L.**

Crtra. Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC1076_24980103/1**

Product	Formic acid, eluent additive for LC-MS	Batch	24980103
AC1076		Quality release date	10/07/2024
		Expiry date	07/2027
Analysis	Batch value	Specifications	
Assay (Acidimetric)	99,36 %	98,0 - 100,0	
Aluminium (Al)	< 0,05 ppm	≤ 0,05	
Barium (Ba)	< 0,05 ppm	≤ 0,05	
Cadmium (Cd)	< 0,05 ppm	≤ 0,05	
Calcium (Ca)	< 0,2 ppm	≤ 0,2	
Chromium (Cr)	< 0,05 ppm	≤ 0,05	
Cobalt (Co)	< 0,05 ppm	≤ 0,05	
Copper (Cu)	< 0,05 ppm	≤ 0,05	
Iron (Fe)	< 0,2 ppm	≤ 0,2	
Lead (Pb)	< 0,05 ppm	≤ 0,05	
Lithium (Li)	< 0,05 ppm	≤ 0,05	
Magnesium (Mg)	< 0,05 ppm	≤ 0,5	
Manganese (Mn)	< 0,05 ppm	≤ 0,05	
Nickel (Ni)	< 0,05 ppm	≤ 0,05	
Molybdenum (Mo)	< 0,05 ppm	≤ 0,05	
Potassium (K)	< 0,1 ppm	≤ 0,1	
Silver (Ag)	< 0,05 ppm	≤ 0,05	
Sodium (Na)	< 0,05 ppm	≤ 0,5	
Strontium (Sr)	< 0,05 ppm	≤ 0,05	
Thallium (Tl)	< 0,05 ppm	≤ 0,05	
Zinc (Zn)	< 0,05 ppm	≤ 0,05	
Suitability for use in LC-MS	passes test	passes test	

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martinez Casals
Laboratory Technician