

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC1083_23229101/1**

Product Formic acid, solution 90,1% ± 0,1% w/w, for analysis, ExpertQ®

Batch 23229101

AC1083

Quality release date 18/10/2022

Expiry date 10/2025

Analysis	Batch value	Specifications
Assay (Acidimetric)	90,20 %	90,0 - 90,2
Colour (Hazen)	5	≤ 10
Acetic acid (CH ₃ COOH)	< 0,05 %	≤ 0,05
Chlorides (Cl)	< 0,0005 %	≤ 0,0005
Sulfates (SO ₄)	< 0,0005 %	≤ 0,0005
Sulfites (SO ₃)	< 0,001 %	≤ 0,001
Aluminium (Al)	< 0,05 ppm	≤ 0,05
Ammonium (NH ₄)	< 0,001 %	≤ 0,001
Barium (Ba)	< 0,05 ppm	≤ 0,05
Beryllium (Be)	< 0,02 ppm	≤ 0,02
Bismuth (Bi)	< 0,1 ppm	≤ 0,1
Cadmium (Cd)	< 0,05 ppm	≤ 0,05
Calcium (Ca)	< 0,2 ppm	≤ 0,2
Chromium (Cr)	< 0,05 ppm	≤ 0,05
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,02 ppm	≤ 0,02
Germanium (Ge)	< 0,05 ppm	≤ 0,05
Heavy metals (as Pb)	< 5 ppm	≤ 5
Iron (Fe)	< 2 ppm	≤ 2
Lead (Pb)	< 0,02 ppm	≤ 0,02
Lithium (Li)	< 0,02 ppm	≤ 0,02
Magnesium (Mg)	< 0,5 ppm	≤ 0,5
Manganese (Mn)	< 0,05 ppm	≤ 0,05
Molybdenum (Mo)	< 0,02 ppm	≤ 0,02
Nickel (Ni)	< 0,05 ppm	≤ 0,05
Potassium (K)	< 0,1 ppm	≤ 0,1
Silver (Ag)	< 0,02 ppm	≤ 0,02
Sodium (Na)	< 0,5 ppm	≤ 0,5
Strontium (Sr)	< 0,02 ppm	≤ 0,02
Thallium (Tl)	< 0,05 ppm	≤ 0,05
Titanium (Ti)	< 0,1 ppm	≤ 0,1
Vanadium (V)	< 0,05 ppm	≤ 0,05
Zinc (Zn)	< 0,05 ppm	≤ 0,05
Zirconium (Zr)	< 0,1 ppm	≤ 0,1
Dilution test	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,001

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

AC1083_23229101/1

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martinez Casals
Laboratory Technician

A handwritten signature in black ink, appearing to be "Marta", written over a horizontal line.