

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

Product	Water, gradient HPLC grade	Batch	18734703
AG0001		Quality release date	27/02/2018
		Expiry date	02/2020
Analysis	Batch value	Specifications	
Conductivity (25°C)	0,8 µS/cm	≤ 1	
Chlorides (Cl)	< 0,00002 %	≤ 0,00002	
Nitrates (NO3)	< 0,00003 %	≤ 0,00003	
Sulfates (SO4)	< 0,0001 %	≤ 0,0001	
Residue on evaporation	0,0001 %	≤ 0,0001	
Lead (Pb)	< 0,00001 %	≤ 0,00001	
Colony count	< 25 UFC/g	≤ 25	
Microbiological test	passes test	passes test	
Gradient elution: maximum absorption of the largest eluted peaks:			
at 210 nm	< 0,01 AU	≤ 0,01	
at 254 nm	< 0,001 AU	≤ 0,001	

Microfiltered through membranes of pore diameter 0,22 µm