

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

Product	Water, gradient HPLC grade	Batch	19460603
AG0001		Quality release date	08/01/2019
		Expiry date	01/2021
Analysis		Batch value	Specifications
Conductivity (25°C)		0,7 µS/cm	≤ 1
Chlorides (Cl)		< 0,00002 %	≤ 0,00002
Nitrates (NO3)		< 0,00003 %	≤ 0,00003
Sulfates (SO4)		< 0,0001 %	≤ 0,0001
Residue on evaporation		0,0001 %	≤ 0,0001
Lead (Pb)		< 0,00001 %	≤ 0,00001
Colony count		< 25 UFC/g	≤ 25
Microbiological test		passes test	passes test
Gradient elution: maximum absorption of the largest eluted peaks:			
at 210 nm		< 0,01 AU	≤ 0,01
at 254 nm		< 0,001 AU	≤ 0,001

Microfiltered through membranes of pore diameter 0,22 µm