

Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2 08181 Sentmenat

Expiry date

1 mg/L

1 mg/L

1 mg/L

1 mg/L

0,4 mg/L

0,4 mg/L

12/2020

L +34937456400

CERTIFICATE OF ANALYSIS AG0014_20626106/1

≤ 1 mg/L

≤ 1 mg/L

≤ 1 mg/L

≤ 1 mg/L

≤ 0,4 mg/L

≤ 0,4 mg/L

Product Water, GC head space grade Batch 20626106

AG0014 Quality release date 06/04/2020

Analysis Batch value Specifications Conductivity (25°C) 0,4 µS/cm ≤ 1 Class 1 according to ICH less than 0,5 ppm. absence of peak absence Class 2 and 3 solvents likely to be present below following limits tert-Butyl methyl ether ≤ 1 mg/L 1 mg/L Acetone ≤ 1 mg/L 1 mg/L Methanol ≤ 1 mg/L 1 mg/L Tetrahydrofuran ≤ 0,7 mg/L 0,7 mg/L

Suitable for residual solvent analysis according to guideline CPMP/ICH/283/95, PhEur and USP.

For laboratory use only.

Ethanol

Acetonitrile 2-Propanol

n-Propanol

1,4-Dioxane

Pyridine

The -