



Scharlab S.L.

Tel. int.: +34-93-7151811

Email: scharlab@scharlab.com

CERTIFICATE OF ANALYSIS

Product:	Aluminium, standard solution 1000 mg/l Al for	Batch	17075901
AL0751	AA (aluminium nitrate nonahydrate in nitric acid 0,5	Quality release date	31/03/2016
	mol/l)	Expiry date	3/2019

Analysis	Batch value	Specifications
concentration	1003 mg/l	995 - 1005 mg/l
uncertainty ± 5 mg/l		

This standard solution is traceable to

Standard Reference Material from NIST.

Preparation

This atomic absorption standard has been prepared using gravimetric and volumetric procedures.

Composition per litre is 13,90 g Aluminium nitrate nonahydrate in Nitric acid 0,5 mol/l

Traceability

This standard is traceable to Standard Reference Material from NIST.

SRM 915b Calcium carbonate

Uncertainty

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the expanded uncertainty are given at a confidence level of 95% ($k=2$).

Measurement

The batch value is determined by direct complexometry titration at 20°C, by means of spectrophotometric-end-point.

Batch value certified at the time of measurement.

Storage and use

For AA spectrophotometer calibration

If the product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.

Once the bottle is opened, keep tightly closed at room temperature. Avoid exposure to light.

We suggest rejecting the solution six months after opening

This solution should be checked periodically by the user.

This certificate does not release the user from their control upon receipt of the goods

You can get a copy of any of our COA from our web site: www.scharlab.com

M. Canet

Laboratory Manager