

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AN0073_25118901/1**

Product Methyl orange, C.I. 13025, indicator, for analysis, ExpertQ®, ACS

Batch 25118901

AN0073

Quality release date 17/09/2024

Expiry date 09/2034

Analysis	Batch value	Specifications
Appearance of solution	passes test	passes test
pH range (pink to orange-yellow)	3,2 - 4,4	3,2 - 4,4
Visual transition interval	passes test	passes test
Absorption maximum λ_1 (pH 3,1)	504 nm	501 - 504
Absorption maximum λ_2 (pH 4,4)	471 nm	467 - 471
Absorptivity (A1%/1 cm; λ_1 ; pH 3,1 on dried sample)	1150	1050 - 1150
Absorptivity (A1%/1 cm; λ_2 ; pH 4,4 on dried sample)	850	750 - 850
Loss on drying (110 °C)	< 5 %	≤ 5

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician