

## Scharlab S.L.

Crtra. Polinya a Sentmenat, Km 8,2 08181 Sentmenat

**L** +34937456400

# CERTIFICATE OF ANALYSIS AQ0001 24831901/2

Product Aquagent® Titrant 5, free from pyridine, titrant-component for volumetric Karl Fischer titration		Batch 248 Quality release date 03/06	
AQ0001		Expiry date 05	5/2027
Analysis	Batch value	Specifications	±U
Factor	4,94	4,9 - 5,1	0,03

### Preparation

This reagent is prepared using gravimetric and volumetric procedures. Contains methanol.

#### Measurament

Batch value certified at the time of measurement.

The equipments involved in the analysis of this reagent are validated periodically.

#### Storage and use

For KF titration.

For use with AQ0029 Aquagent® Solvent.

If the product is stored and unopened, this reagent is stable for 3 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.

For laboratory use only

