

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0002_20605701/1**

Product Aquagent® Solvent, free from pyridine, solvent-component for volumetric Karl Fischer titration

Batch 20605701

Quality release date 27/02/2020

Expiry date 02/2025

AQ0002**Analysis**

Density (20°/4°)

Suitability for volumetric KF titration

Batch value

0,9100 g/mL

passes test

Specifications

0,89 - 0,91

passes test

Preparation

This reagent is prepared using gravimetric and volumetric procedures.
Contains Imidazole, sulfur dioxide and methanol.

Measurement

Batch value certified at the time of measurement.

The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.

For use with AQ0001 Aquagent® Titrant 5 or AQ0006 Aquagent® Titrant 2.

If the product is stored and unopened, this reagent is stable for 5 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature.

For laboratory use only