

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AQ0004\_23738401/2**

Product	Aquagent® Complet 5K, free from pyridine, one-component reagent for volumetric Karl Fischer titration (ketones, aldehydes)		Batch	23738401
			Quality release date	20/04/2023
<b>AQ0004</b>			Expiry date	04/2025
<b>Analysis</b>	<b>Batch value</b>	<b>Specifications</b>	<b>±U</b>	
Factor	4,97	4,5 - 5,5	0,02	

**Preparation**

This reagent is prepared using gravimetric and volumetric procedures.  
Contains imidazole, sulfur dioxide and diethyleneglycol monoethyl ether.

**Measurement**

Batch value certified at the time of measurement.  
The equipments involved in the analysis of this reagent are validated periodically.

**Storage and use**

For KF titration.  
If the product is stored and unopened, this reagent is stable for 2 years from the date of manufacturing.  
Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.

For laboratory use only