

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0005_23966401/1**

Product	Aquagent® Medium K, free from pyridine, solvent for volumetric Karl Fischer titration (ketones, aldehydes)	Batch	23966401
		Quality release date	05/07/2023
AQ0005		Expiry date	07/2025

Analysis	Batch value	Specifications
Suitability for volumetric KF titration	passes test	passes test

Preparation

This reagent is prepared using gravimetric and volumetric procedures.
Contains Trichloromethane and 2-Chloroethanol

Measurement

Batch value certified at the time of measurement.
The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.
For use with AQ0004 Aquagent® Complet 5K.
If the product is stored and unopened, this reagent is stable for 2 years from the date of manufacturing.
Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature.

For laboratory use only