

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0006_23893601/2**

Product AQ0006	Aquagent® Titrant 2, free from pyridine, titrant-component for volumetric Karl Fischer titration		Batch	23893601
			Quality release date	08/06/2023
			Expiry date	06/2026
Analysis	Batch value	Specifications	±U	
Factor	2,04	1,90 - 2,20	0,02	

Preparation

This reagent is prepared using gravimetric and volumetric procedures.
Contains methanol.

Measurement

Batch value certified at the time of measurement.
The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.
For use with AQ0002 Aquagent® Solvent.
If the product is stored and unopened, this reagent is stable for 3 years from the date of manufacturing.
Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.
For laboratory use only