

Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2 08181 Sentmenat

**** +34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS AQ0006_24590701/1

Product Aquagent® Titrant 2, free from pyridine, titrant-component for volumetric Karl Fischer titration AQ0006		Batch Quality release date Expiry date	24590701 29/02/2024 02/2027	
Analysis	Batch value	Specification	S	±U
Factor	1,98	1,90 - 2,20	(0,01

Preparation

This reagent is prepared using gravimetric and volumetric procedures. Contains methanol.

Measurement

Batch value certified at the time of measurement.

The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.

For use with AQ0002 Aquagent® Solvent.

If the product is stored and unopened, this reagent is stable for 3 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.

For laboratory use only

