

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AQ0011\_22347801/1**

Product      Aquagent® Metanol Fast  
**AQ0011**

Batch      22347801  
Quality release date      16/12/2021  
Expiry date      12/2026

**Analysis**

Suitability for volumetric KF titration

**Batch value**

passes test

**Specifications**

passes test

**Preparation**

This reagent is prepared using gravimetric and volumetric procedures.  
Contains imidazole, sulfur dioxide and methanol.

**Storage and use**

For KF volumetric titration. Reagent for accelerated volumetric titration.

For use with Aquagent® complet 5 (AQ0003) or Aquagent® complet 2 (AQ0007).

If the product is stored and unopened, this reagent is stable for 5 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature.

For laboratory use only