

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0011_22602501/1**

Product Aquagent® Metanol Fast
AQ0011

Batch 22602501
Quality release date 15/03/2022
Expiry date 03/2027

Analysis

Suitability for volumetric KF titration

Batch value

passes test

Specifications

passes test

Preparation

This reagent is prepared using gravimetric and volumetric procedures.
Contains imidazole, sulfur dioxide and methanol.

Storage and use

For KF volumetric titration. Reagent for accelerated volumetric titration.

For use with Aquagent® complet 5 (AQ0003) or Aquagent® complet 2 (AQ0007).

If the product is stored and unopened, this reagent is stable for 5 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature.

For laboratory use only