

**Scharlab S.L.**

Crtra. Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AQ0014\_24935501/1**

Product	Aquagent Coulometric CG	Batch	24935501
<b>AQ0014</b>		Quality release date	26/06/2024
		Expiry date	06/2029

Analysis	Batch value	Specifications
Suitability for coulometric KF titration	passes test	passes test
Water content	32,6 ppm	≤ 50

**Composition**

Contains methanol.

**Storage and use**

Cathode solution for coulometric Karl-Fischer titration.

It has a water capacity of approx. 300 mg per 5 mL.

Suitable for cells with diaphragm.

If the product is stored and unopened, this reagent is stable for 3 years from the date of manufacturing.

Protect from moisture. Keep tightly closed at room temperature.

For laboratory use only