

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0014_25643401/2**

Product	Aquagent Coulometric CG	Batch	25643401
AQ0014		Quality release date	13/03/2025
		Expiry date	03/2030

Analysis	Batch value	Specifications
Suitability for coulometric KF titration	passes test	passes test
Water content	40,6 ppm	≤ 50

Composition

Contains methanol.

Storage and use

Cathode solution for coulometric Karl-Fischer titration.

It has a water capacity of approx. 300 mg per 5 mL.

Suitable for cells with diaphragm.

If the product is stored and unopened, this reagent is stable for 5 years from the date of manufacturing.

Protect from moisture. Keep tightly closed at room temperature.

For laboratory use only