

### Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa 08181 Sentmenat

**L** +34937456400

☑ helpdesk@scharlab.com

# CERTIFICATE OF ANALYSIS AQ0020 25430401/2

Product AQ0020	Aquagent®, standard solution 1%		Batch 2 Quality release date 15 Expiry date	25430401 5/01/2025 01/2028
Analysis		Batch value	Specifications	±U
Water content		10,05 mg/g	9,85 - 10,15	0,04
Water content		1,005 %	0,985 - 1,015	0,004

# **Preparation**

This standard is prepared using gravimetric and volumetric procedures. Contains 1-butanol, propylene carbonate and xylene.

#### Measurement

The water content is determined by means of direct volumetric Karl-Fischer titration under ISO/IEC 17025. The stated value is obtained as an average of fifteen measurements at five distinct samples. The standard is traceable to SI-units (kg) and verified against a standard traceable to NIST SRM 2890 (Water Saturated 1-Octanol).

## **Uncertainty**

The uncertainty of the analytical method is calculated using the Kragten method and in accordance with GUM. The limits of the uncertainty are given at a confidence level of 95% (k=2).

## Storage and use

For Karl-Fischer titrator calibration.

If the product is properly stored and unopened, this standard is stable for 3 years from the date of manufacturing.

The standard must be kept tightly closed at room temperature. Avoid exposure to light and moisture.

For laboratory use only.

