

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0021_22348401/1**

Product Aquagent®, standard solution 5
AQ0021

Batch 22348401
Quality release date 21/12/2021
Expiry date 12/2024

Analysis	Batch value	Specifications
Water content	5,00 mg/mL	4,98 - 5,02
Water content	5,88 mg/g	5,86 - 5,90

Preparation

This standard is prepared using gravimetric and volumetric procedures.
Contains 1-butanol and xylene.

Measurement

Batch value certified at the time of measurement.
The equipments involved in the analysis of this standard are validated periodically.

Storage and use

For KF equipment calibration.
If the product is stored and unopened, this standard is stable for 3 years from the date of manufacturing.
The standard is bottled in an anhydrous atmosphere. Protect from moisture.
Shake gently the bottle before use.
Samples must be withdrawn from the bottle into a dry syringe or glassware.
The standard must be kept tightly closed in a refrigerator once opened.
For laboratory use only