

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AQ0021\_25832701/1**

Product	Aquagent®, standard solution 0,5%	Batch	25832701
<b>AQ0021</b>		Quality release date	19/05/2025
		Expiry date	05/2028
<b>Analysis</b>	<b>Batch value</b>	<b>Specifications</b>	
Water content	4,99 mg/mL	4,98 - 5,02	
Water content	5,88 mg/g	5,86 - 5,90	
Water content	0,499 %	0,498 - 0,502	

**Preparation**

This standard is prepared using gravimetric and volumetric procedures.  
Contains 1-butanol and xylene.

**Measurement**

Batch value certified at the time of measurement.  
The equipments involved in the analysis of this standard are validated periodically.

**Storage and use**

For KF equipment calibration.  
If the product is stored and unopened, this standard is stable for 3 years from the date of manufacturing.  
The standard is bottled in an anhydrous atmosphere. Protect from moisture.  
Shake gently the bottle before use.  
Samples must be withdrawn from the bottle into a dry syringe or glassware.  
The standard must be kept tightly closed in a refrigerator once opened.

For laboratory use only