

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0029_24471401/1**

Product	Aquagent® Solvent, free from pyridine, solvent for volumetric two-components Karl Fischer titration	Batch	24471401
		Quality release date	23/01/2024
AQ0029		Expiry date	01/2029

Analysis	Batch value	Specifications
Density (20°/4°)	0,8927 g/mL	0,85 - 0,91
Suitability for volumetric KF titration	passes test	passes test
Water content	100,0 ppm	100 - 200

Preparation

This reagent is prepared using gravimetric and volumetric procedures.
Contains Imidazole, sulfur dioxide and methanol.

Measurement

Batch value certified at the time of measurement.
The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.
For use with AQ0001 Aquagent® Titrant 5 or AQ0006 Aquagent® Titrant 2.
If the product is stored and unopened, this reagent is stable for 5 years from the date of manufacturing.
Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature.
For laboratory use only.