

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AQ0029\_25986001/1**

---

Product	Aquagent® Solvent, free from pyridine, solvent for volumetric two-components Karl Fischer titration	Batch	25986001
		Quality release date	14/07/2025
<b>AQ0029</b>		Expiry date	07/2030

---

Analysis	Batch value	Specifications
Suitability for volumetric KF titration	passes test	passes test
Water content	298 ppm	100 - 300

**Preparation**

This reagent is prepared using gravimetric and volumetric procedures.  
Contains Imidazole, sulfur dioxide and methanol.

**Measurement**

Batch value certified at the time of measurement.  
The equipments involved in the analysis of this reagent are validated periodically.

**Storage and use**

For KF titration.  
For use with AQ0001 Aquagent® Titrant 5 or AQ0006 Aquagent® Titrant 2.  
If the product is stored and unopened, this reagent is stable for 5 years from the date of manufacturing.  
Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature.  
For laboratory use only.