

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0030_23168602/1**

| | | | |
|---------------|---|----------------------|------------|
| Product | Aquagent®, di-Sodium tartrate dihydrate, secondary standard for volumetric Karl-Fischer titration | Batch | 23168602 |
| | | Quality release date | 28/09/2022 |
| AQ0030 | | Expiry date | 09/2027 |

| Analysis | Batch value | Specifications |
|----------|-------------|----------------|
| Water | 15,70 % | 15,61 - 15,71 |

Measurement

The batch value is determined by means of Karl-Fischer titration against a Scharlau's standard solution 10.

Uncertainty

The uncertainty of the analytical method is $\pm 0,05\%$

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the uncertainty are given at a confidence level of 95% ($k=2$).

Storage and use

For Karl-Fischer titrator calibration.

If the product is stored and unopened, this standard is stable for 5 years from the date of manufacturing.

The standard must be kept tightly closed at room temperature. Avoid exposure to light and moisture.

For laboratory use only