

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0033_26157301/2**

Product	Aquagent® Complet 2, free from pyridine, one-component reagent for volumetric Karl Fischer titration	Batch	26157301
		Quality release date	06/10/2025
		Expiry date	10/2028
AQ0033			
Analysis	Batch value	Specifications	±U
Factor	2,12	1,60 - 2,40	0,05

Preparation

This reagent is prepared using gravimetric and volumetric procedures.

Contains imidazole, 2-methylimidazole, sulfur dioxide and diethyleneglycol monoethyl ether.

Measurement

Batch value certified at the time of measurement.

The titer has been determined according to ISO 760 and ASTM E203.

The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.

If the product is stored and unopened, this reagent is stable for 3 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.

For laboratory use only