

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0034_26275601/2**

Product	Aquagent® Complet 5K, free from pyridine, one-component reagent for volumetric Karl Fischer titration (Ketone & Aldehyde)	Batch	26275601
		Quality release date	17/11/2025
AQ0034		Expiry date	11/2028

Analysis	Batch value	Specifications	±U
Factor	5,33	4,5 - 5,5	0,05

Preparation

This reagent is prepared using gravimetric and volumetric procedures.

Contains imidazole, 2-methylimidazole, sulfur dioxide and diethyleneglycol monoethyl ether.

Measurement

Batch value certified at the time of measurement.

The titer has been determined according to ISO 760 and ASTM E203.

The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.

If the product is stored and unopened, this reagent is stable for 3 years from the date of manufacturing.

Protect from moisture.

For laboratory use only