



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

Product	Bromophenol blue, indicator, ACS	Batch	18743901
AZ0125		Quality release date	28/02/2018
		Expiry date	02/2028
Analysis	Batch value	Specifications	
pH range (greenish-yellow to blue-violet)	3,1 - 4,4	3,1 - 4,4	
Absorption maximum I1 (pH 3,0)	438 nm	434 - 439	
Absorption maximum I2 (pH 4,6)	591 nm	590 - 593	
Absorptivity (A1%/1 cm; I1; pH 3,0 on dried sample)	385	350 - 385	
Absorptivity (A1%/1 cm; I2; pH 4,6 on dried sample)	1000	940 - 1000	
Transition range acc. ACS	passes test	passes test	
Loss on drying (110 °C)	0,8 %	≤ 1	

This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

M. Canet
Laboratory Manager