



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

Product	Bromothymol blue, indicator, ACS	Batch	19968801
AZ0130		Quality release date	03/07/2019
		Expiry date	06/2029
Analysis	Batch value	Specifications	
pH range (yellow to blue)	5,8 - 7,6	5,8 - 7,6	
Absorption maximum λ_1 (pH 5,8)	430 nm	430 - 435	
Absorption maximum λ_2 (pH 7,6)	616 nm	615 - 618	
Absorptivity (A1%/1 cm; λ_1 ; pH 5,8 on dried sample)	300	260 - 300	
Absorptivity (A1%/1 cm; λ_2 ; pH 7,6 on dried sample)	470	470 - 520	
Transition range acc. ACS	passes test	passes test	
Loss on drying (110 °C)	1,83 %	≤ 3	