

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**CL0345_20997712/1**

Product Dichloromethane, for GC residue analysis, stabilized with approx. 50 ppm of amylene

Batch 20997712

Quality release date 14/07/2020

Expiry date 07/2023

CL0345

| Analysis | Batch value | Specifications |
|------------------------|-------------|----------------|
| Assay (G.C.) | 99,99 % | ≥ 99,9 |
| Identity (IR-Spectrum) | passes test | passes test |
| Density (20°/4°) | 1,3267 g/mL | 1,323 - 1,327 |
| Residue on evaporation | 0,0001 % | ≤ 0,0001 |
| Water (K.F.) | 0,0092 % | ≤ 0,02 |

Suitable for organohalogenated pesticide and dioxins, furans and PCBs residue analysis.

ECD, from 1,2,4-trichlorobenzene to decachlorobiphenyl, no peaks are obtained greater than 3 pg/ml as lindane.

No peaks are obtained in vicinity of 2,4,5-trichlorobiphenyl.

Suitable for pesticide and polycyclic aromatic hydrocarbons residue analysis.

FID, from 1-octanol to 1-tetradecanol, no peaks are obtained greater than 5 ng/ml as 1-tetradecanol.

No peaks are obtained in vicinity of pyrene.

For laboratory use only.