

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****CL0346\_24236108/1**

Product Dichloromethane, GC-MS, suitable for nitrosamine analysis

Batch 24236108

**CL0346**

Quality release date 25/10/2023

Expiry date 10/2026

Analysis	Batch value	Specifications
Assay (G.C.)	99,86 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Colour (Hazen)	4	≤ 10
Residue on evaporation	0,9 ppm	≤ 3
Water (K.F.)	0,0066 %	≤ 0,05
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0
Nitrosamine analysis	passes test	passes test

Suitable for residue analysis.

For laboratory use only.