

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**CL0346_26255908/1**

Product	Dichloromethane, GC-MS, suitable for nitrosamine analysis	Batch	26255908
CL0346		Quality release date	06/11/2025
		Expiry date	11/2028

Analysis	Batch value	Specifications
Assay (G.C.)	99,88 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Colour (Hazen)	4	≤ 10
Residue on evaporation	0,9 ppm	≤ 3
Water (K.F.)	0,005 %	≤ 0,05
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0
Nitrosamine analysis	passes test	passes test

Suitable for residue analysis.

For laboratory use only.