

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**CO0085_25196001/1**

Product	Copper, standard solution 1000 mg/l Cu for AA (copper(II) nitrate in nitric acid 0,5 mol/l)	Batch	25196001
		Quality release date	08/10/2024
CO0085		Expiry date	10/2027

Analysis	Batch value	Specifications
Concentration	1001,5 mg/L	995,0 - 1005,0

Preparation

This atomic absorption standard has been prepared using gravimetric and volumetric procedures.
Composition per litre is 3,80 g Copper (II) nitrate trihydrate in Nitric acid 0,5 mol/l.

Traceability

This standard is traceable to Standard Reference Material from NIST.
SRM 915b Calcium carbonate

Uncertainty

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the expanded uncertainty are given at a confidence level of 95% ($k=2$).

Measurement

The batch value is determined by direct complexometry titration at 20°C, by means of spectrophotometric-end-point.
Batch value certified at the time of measurement.

Storage and use

For AA spectrophotometer calibration.
If the product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.
Once the bottle is opened, keep tightly closed at room temperature. Avoid exposure to light.
We suggest rejecting the solution six months after opening.
This solution should be checked periodically by the user.

For laboratory use only