

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**DI1115_23362501/1**

Product Dimidium bromide, for determination of tensioactives

Batch 23362501

DI1115

Quality release date 01/12/2022

Expiry date 11/2027

Analysis	Batch value	Specifications
Assay (titration with HClO ₄)	98,00 %	≥ 98
Absorptivity (A1%/1 cm; λ max; methanol)	175	155 - 175
Absorption maximum (in methanol)	523 nm	520 - 523
Loss on drying (110 °C)	< 5 %	≤ 5

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martinez Casals
Laboratory Technician