



## Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

# CERTIFICATE OF ANALYSIS

### DT0004\_26159601/2

Product Deterlabo® N, neutral detergent, concentrated liquid for manual use

Batch 26159601

**DT0004**

Quality release date 02/10/2025

Expiry date 10/2028

Analysis	Batch value	Specifications
Appearance	clear and bright blue	clear and bright blue
Density (20°/20°)	1,0410 g/mL	1,01 - 1,05
pH	7,12	6,5 - 7,5

#### Contains:

Anionic surfactants between 5 - 15 %.

Non-ionic surfactants < 5 %.

Perfumers.

Allergens: Limonene.

Preservatives: Benzyl isothiazolone; Methyl isothiazolinone.

#### Dosage:

Dilute in water between 2 - 15 g/L.

#### Properties and applications:

Neutral detergent, concentrated liquid of high foaming power.

Best-suited for manual cleaning of all types of equipment, tools, material and laboratory surfaces.

It can be used for cleaning stainless steel, aluminium, glass, formica, vinyl, painted or varnished surfaces, crystal, pyrex, etc.

For laboratory use only

This certificate does not exempt the user from checking the results upon receipt of the goods.  
Any copy of our CoA may be obtained from our website at [www.scharlab.com](http://www.scharlab.com)  
This certificate modifies and cancels the content of the certificate with name S-25-8739/1

Francisco Javier Ruiz Baena  
Laboratory Coordinator