



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ET0004_22626603/1

Product Ethanol 96% v/v, for analysis, ExpertQ®, ACS, Reag. Ph Eur

Batch 22626603

ET0004

Quality release date 23/03/2022

Expiry date 03/2027

Analysis	Batch value	Specifications
Assay (G.C.) (v/v)	96,17 %	95,1 - 96,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,8040 g/mL	0,804 - 0,807
Appearance	clear	clear
Colour (Hazen)	2	≤ 10
Solubility in water	passes test	passes test
Acidity	0,00010 meq/g	≤ 0,0002
Alkalinity	0,00005 meq/g	≤ 0,0002
Chlorides (Cl)	< 0,00003 %	≤ 0,00003
Nitrates (NO ₃)	< 0,00003 %	≤ 0,00003
Phosphates (as PO ₄)	< 0,00003 %	≤ 0,00003
Sulfates (SO ₄)	< 0,00003 %	≤ 0,00003
Aluminium (Al)	< 0,5 ppm	≤ 0,5
Antimony (Sb)	< 0,02 ppm	≤ 0,02
Arsenic (As)	< 0,02 ppm	≤ 0,02
Barium (Ba)	< 0,1 ppm	≤ 0,1
Beryllium (Be)	< 0,02 ppm	≤ 0,02
Bismuth (Bi)	< 0,02 ppm	≤ 0,02
Boron (B)	< 0,02 ppm	≤ 0,02
Cadmium (Cd)	< 0,05 ppm	≤ 0,05
Calcium (Ca)	< 0,5 ppm	≤ 0,5
Chromium (Cr)	< 0,02 ppm	≤ 0,02
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,02 ppm	≤ 0,02
Gallium (Ga)	< 0,02 ppm	≤ 0,02
Gold (Au)	< 0,02 ppm	≤ 0,02
Indium (In)	< 0,02 ppm	≤ 0,02
Iron (Fe)	< 0,1 ppm	≤ 0,1
Lead (Pb)	< 0,1 ppm	≤ 0,1
Lithium (Li)	< 0,02 ppm	≤ 0,02
Magnesium (Mg)	< 0,1 ppm	≤ 0,1
Manganese (Mn)	< 0,02 ppm	≤ 0,02
Molybdenum (Mo)	< 0,02 ppm	≤ 0,02
Nickel (Ni)	< 0,02 ppm	≤ 0,02
Platinum (Pt) ppm	< 0,02 ppm	≤ 0,02
Silver (Ag)	< 0,02 ppm	≤ 0,02
Thallium (Tl)	< 0,02 ppm	≤ 0,02
Tin (Sn)	< 0,1 ppm	≤ 0,1
Vanadium (V)	< 0,02 ppm	≤ 0,02
Zinc (Zn)	< 0,1 ppm	≤ 0,1
Zirconium (Zr)	< 0,02 ppm	≤ 0,02

Results marked with # are out of specifications.

This certificate does not exempt the user from checking the results upon receipt of the goods.

Any copy of our CoA may be obtained from our website at www.scharlab.com

Sara Garcia Maldonado
Laboratory Technician

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ET0004_22626603/1**

Product Ethanol 96% v/v, for analysis, ExpertQ®, ACS, Reag. Ph Eur

Batch 22626603

ET0004

Quality release date 23/03/2022

Expiry date 03/2027

Analysis	Batch value	Specifications
Formaldehyde	< 0,0005 %	≤ 0,0005
Furfural	passes test	passes test
Fusel oil	passes test	passes test
Acetaldehyde and acetal (G.C.)	< 0,001 %	≤ 0,001
Acetone (G.C.)	< 0,001 %	≤ 0,001
Benzene (G.C.)	< 0,0002 %	≤ 0,0002
Isoamyl alcohol (G.C.)	< 0,05 %	≤ 0,05
Methanol (G.C.)	< 0,01 %	≤ 0,01
Methylethylketone (G.C.)	< 0,02 %	≤ 0,02
2-Propanol (G.C.)	< 0,003 %	≤ 0,003
Aldehydes (as CH ₃ CHO)	< 0,001 %	≤ 0,001
Carbonyl compounds (as CO)	< 0,003 %	≤ 0,003
Higher alcohols (G.C.)	< 0,01 %	≤ 0,01
Substances darkened by H ₂ SO ₄	passes test	passes test
Substances reducing KMnO ₄	passes test	passes test
Residue on evaporation	0,0002 %	≤ 0,0005
Water (v/v) (K.F.)	3,83 %	3,1 - 4,9

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Sara Garcia Maldonado
Laboratory Technician