

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ET0010_25554106/1**

Product	Ethanol absolute, gradient HPLC grade	Batch	25554106
ET0010		Quality release date	14/02/2025
		Expiry date	02/2030
Analysis	Batch value	Specifications	
Assay (G.C.) (v/v)	99,99 %	≥ 99,9	
Identity (IR-Spectrum)	passes test	passes test	
Density (20°/4°)	0,7894 g/mL	0,789 - 0,790	
Acidity	0,00014 meq/g	≤ 0,0002	
Alkalinity	0,00001 meq/g	≤ 0,0002	
Residue on evaporation	0,0002 %	≤ 0,0002	
Water (v/v) (K.F.)	0,0578 %	≤ 0,1	
Gradient grade (254 nm):			
maximum background absorbance:	< 0,02 AU	≤ 0,02	
maximum peak absorbance:	< 0,002 AU	≤ 0,002	
UV Spectroscopy in a 1,0 cm cell:			
min. transmission at 205 nm	20,00 %	≥ 20	
min. transmission at 220 nm	63,63 %	≥ 50	
min. transmission at 245 nm	93,80 %	≥ 90	
min. transmission at 260 nm	98,55 %	≥ 98	
max. absorbance at 205 nm	0,699 AU	≤ 0,699	
max. absorbance at 220 nm	0,196 AU	≤ 0,301	
max. absorbance at 245 nm	0,028 AU	≤ 0,046	
max. absorbance at 260 nm	0,006 AU	≤ 0,009	

Microfiltered through membranes of pore diameter 0,22 µm.

Suitable for UPLC.

For laboratory use only