

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

Product	Ethanol absolute, HPLC BASIC	Batch	20613215
ET0024		Quality release date	03/03/2020
		Expiry date	03/2025
Analysis	Batch value	Specifications	
Assay (G.C.) (v/v)	99,99 %	≥ 99,9	
Acidity	0,00013 meq/g	≤ 0,0003	
Alkalinity	0,00001 meq/g	≤ 0,0002	
Acetone (G.C.)	< 0,0005 %	≤ 0,001	
Methanol (G.C.)	0,0001 %	≤ 0,05	
2-Propanol (G.C.)	0,0002 %	≤ 0,003	
Substances darkened by H ₂ SO ₄	passes test	passes test	
Substances reducing KMnO ₄	passes test	passes test	
Residue on evaporation	0,0002 %	≤ 0,0005	
Water (v/v) (K.F.)	0,0265 %	≤ 0,1	
Absorbance	passes test	passes test	

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Ricard Balladares Abella
Laboratory Technician