

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**HE0238_25541610/1**

Product	n-Hexane, 96%, for GC residue analysis	Batch	25541610
HE0238		Quality release date	11/02/2025
		Expiry date	02/2028

Analysis	Batch value	Specifications
Assay (G.C.)	97,10 %	≥ 96
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,6601 g/mL	0,659 - 0,662
Residue on evaporation	0,0001 %	≤ 0,0001
Water (K.F.)	0,0021 %	≤ 0,01

Suitable for organohalogenated pesticide and dioxins, furans and PCBs residue analysis.
ECD, from 1,2,4-trichlorobenzene to decachlorobiphenyl, no peaks are obtained greater than 3 pg/ml as lindane.
No peaks are obtained in vicinity of 2,4,5-trichlorobiphenyl.

For laboratory use only.