

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****HE0248\_23906913/1**

Product	n-Hexane, GC-MS	Batch	23906913
<b>HE0248</b>		Quality release date	12/06/2023
		Expiry date	06/2026
<b>Analysis</b>	<b>Batch value</b>	<b>Specifications</b>	
Assay (G.C.)	98,02 %	≥ 96,0	
Identity (IR-Spectrum)	passes test	passes test	
Colour (Hazen)	1	≤ 10	
Residue on evaporation	1,4 ppm	≤ 3	
Water (K.F.)	0,0026 %	≤ 0,05	
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0	

Suitable for residue analysis. For laboratory use only