

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****HE0248\_26208213/1**

---

Product	n-Hexane, GC-MS	Batch	26208213
<b>HE0248</b>		Quality release date	21/10/2025
		Expiry date	10/2028

---

<b>Analysis</b>	<b>Batch value</b>	<b>Specifications</b>
Assay (G.C.)	99,86 %	≥ 96,0
Identity (IR-Spectrum)	passes test	passes test
Colour (Hazen)	5	≤ 10
Residue on evaporation	1,2 ppm	≤ 3
Water (K.F.)	0,002 %	≤ 0,05
GC/MSD (retention range n-undecane to n-tetracontane, scanning area 30 - 600 amu, individual signals (n-tetradecane standard))	< 3,0 ng/mL	≤ 3,0

Suitable for residue analysis.

For laboratory use only

Results marked with # are out of specifications.  
This certificate does not exempt the user from checking the results upon receipt of the goods.  
Any copy of our CoA may be obtained from our website at [www.scharlab.com](http://www.scharlab.com)

Marta Martínez Casals  
Laboratory Technician