

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**IM0026_25853102/1**

Product Imidazole, for analysis, ExpertQ®, ACS

Batch 25853102

IM0026

Quality release date 23/05/2025

Expiry date 05/2030

Analysis	Batch value	Specifications
Assay (titration with HClO ₄)	100,23 %	≥ 99,5
Identity (IR-Spectrum)	passes test	passes test
pH (5 %, H ₂ O)	10,37	9,5 - 11,0
Iron (Fe)	< 5 ppm	≤ 5
Residue on ignition	< 0,1 %	≤ 0,1
Water (K.F.)	0,01 %	≤ 0,2

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician