

# Scharlab S.L.

Gato Pèrez, 33. Pol.Ind. Mas d'en Cisa 08181 Sentmenat Ĉ +34937456400 ⊠ helpdesk@scharlab.com

# **CERTIFICATE OF ANALYSIS**

Product: Standard Solution of: Mercury (Hg)

Batch 18835201

concentration 1.000 g/l in 10% Nitric Acid

(HNO3)

ME0116 Quality Release Date 27.04.2018

Expiry Date: 04.2021

Analysis Batch Value (mg/l) Specifications (mg/l)

concentration (Hg )  $1002.9 \pm 5.5$  (y) 1000

Density: 1.071 g/cm3 at 20 °C

### **Preparation**

This certified reference material is produced in a clean room, using a highest purity starting material, acid from sub-boiling and  $0.055~\mu\text{S/cm}$  deionized water. The low-density polyethylene bottle was decontaminated by leaching with  $0.055~\mu\text{S/cm}$  deionized water and triple rinse.

The instructions of ISO 17034 were considered for the preparation of this solution.

Contains: HgO 99.999%

## **Traceability**

This standard is traceable to NIST SRM No 3133 Lot 061204; NIST SRM No 3168a Lot 120629 The certified value was obtained by a weighted mean of the results of two independent methods among: Classical Volumetric, Primary Gravimetric, Instrumental (ICP/OES, ICP/MS or IC) according to calibration procedure (y) WQP 5.15.1.24

The calibration curve is drawn using a series of standard solutions prepared from a certified reference material traceable to NIST (SRM) and accredited by laboratories/producers in compliance with ISO/IEC 17025 and/or ISO 17034.

# Uncertainty

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor K=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02

#### Measurement

Batch value certified at the time of measurement.

The certified value is calculated by means of both gravimetric preparation and ICP-OES analysis.

#### Hazardous

The normal laboratory safety precautions should be observed when working with this standard. Please refer to Safety Data Sheet (SDS) to further details.

# Homogeneity

This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. To ensure sufficient homogeneity of the sample prior to use, mix thoroughly by shaking.

### Storage and use

For ICP spectrometer calibration.

If stored unopened in the original packaging, this solution is stable for 3 years from the release date. Shelf life is also limited by the effect of transpiration of solvent through the unopened bottle walls at an average of <0.1% per year. Once the bottle is opened, keep tightly closed at room temperature in the original packaging. Do not pipette directly from the bottle. Do not pour the used solution back in the bottle.

This standard can be used directly or can be diluted in an appropriate high-purity matrix.

Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of standard concentration and the volume used for dilution and divided into the final volume used for dilution.

We recommend that the material used be leached with acids.

We suggest rejecting the solution six months after opening.

This document is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31, ISO Guide 35 and Eurachem/CITAC Guides.

The product is produced by laboratory accredited to ISO 17034 and ISO/IEC 17025

Signature: 🏅

(M. Canet

This certificate does not release the user from their control upon receipt of the goods. You can get a copy of any of our COA from our website: www.scharlab.com

## Trace impurities in the actual solution reported in ppm:

(all values below are nominal and not certified)

all values below are norminal and not certified)					
Ag	<0.0038		Cu	<0.0009	
Al	<0.0018		Dy	<0.0054	
As	<0.016		Er	< 0.0035	
Au	<0.016		Eu	< 0.0039	
В	<0.0078		Fe	0.01	
Ва	< 0.0005		Ga	<0.020	
Ве	< 0.0001		Gd	<0.0028	
Bi	<0.016		Ge	<0.020	
Ca	0.194		Hf	< 0.0032	
Cd	< 0.0012		Hg	*	
Се	<0.0085		Ю	< 0.0053	
Co	<0.0028		In	<0.098	
Cr	< 0.0014		lr	< 0.0061	
Cs	< 0.05		K	< 0.0095	

La	<0.0024
Li	<0.0001
Lu	<0.0062
Mg	0.006
Mn	<0.001
Мо	<0.0024
Na	0.01
Nb	<0.0066
Nd	<0.0058
Ni	0.031
Р	<0.048
Pb	<0.021
Pd	< 0.033
Pr	<0.0046

< 0.0097
< 0.063
<0.0081
<0.0038
<0.0089
< 0.071
<0.020
<0.0016
< 0.023
< 0.037
<0.0058
< 0.050
<0.00006
<0.004

Tb	<0.022
Te	<0.031
Th	<0.014
Ti	<0.0012
Ξ	<0.028
Tm	< 0.0023
$\Box$	<0.45
V	<0.0018
W	<0.017
Υ	<0.0007
Yb	< 0.0003
Zn	< 0.0032
Zr	< 0.0007